

UK NEQAS FOR H&I SCHEME 1A - HLA TYPING PHENOTYPING

ASSESSMENT TYPING RESULTS OF SAMPLE 1A06/2014

DESPATCHED ON 3RD JUNE

CONSENSUS HLA TYPE: HLA-A3, A25; B7, B62; Cw7, -; DR103, DR15; DQ6, DQ7

Lab No.	Assessment					HLA-A	A	B	B	C	C	DR	DR	DQ	DQ	Date received	Date tested
	HLA-A	HLA-B	HLA-Cw	HLA-DR	HLA-DQ												
9	YES	YES		YES	YES	3	25	7	62	7		103	15	6	7	04-Jun	05-Jun
11	YES	YES		YES	YES	3	25 (10)	7	62 (15)	NT	NT	103	15 (2)	6 (1)	7 (3)	04-Jun	05-Jun
20	YES	YES		YES	YES	3	25	7	62	7		103	15	6	7	04-Jun	05-Jun
25	YES	YES		YES		3	25	7	62	NT	NT	103	15	6	7	05-Jun	05-Jun
28	YES	YES		YES	YES	3	25	7	62	NT	NT	103	15	6	7	04-Jun	04 - 10-Jun
41	YES	YES		YES	YES	3	25	7	62	NT	NT	103	15	6	7	04-Jun	09-Jun
54	YES	YES		YES	YES	3	25	7	62	NT	NT	103	15	6	7	04-Jun	05-Jun
58	YES	YES	YES	YES	YES	3	25	7	62	7		103	15	6	7	04-Jun	04-Jun
62	YES	YES		YES	YES	3	25	7	62	NT	NT	103	15	6	7	05-Jun	05-Jun
116	YES	YES				NT	NT	NT	NT	NT	NT	NT	NT	NT	NT		
119	YES	YES	YES	YES	YES	3	25	7	62	7		103	15	6	7	04-Jun	05-Jun
120	YES	YES		YES	YES	3	25	7	62	NT	NT	103	15	6	7	04-Jun	05-Jun
123	YES	YES		YES	YES	3	25	7	62	NT	NT	103	15	6	7	04-Jun	05-Jun
133	YES	YES	YES	YES	YES	3	25	7	62 (15)	7		103	15	6	7 (3)	06-Jun	06-Jun
139	YES	YES		YES	YES	3	25	7	62	NT	NT	103	15	6	7	05-Jun	06-Jun
142	YES	YES		YES	YES	3	25	7	62	NT	NT	103	15	6	7	04-Jun	05-Jun
145	YES	YES	YES	YES	YES	3	25	7	62	7		103	15	6	7	04-Jun	05-Jun, 13-Jun
149	YES	YES	YES			3	25	7	62	7		NT	NT	NT	NT	05-Jun	06-Jun
154	YES	YES		YES	YES	3	25	7	62	NT	NT	103	15	6	7	04-Jun	04-Jun
159	YES	YES		YES	YES	3	25	7	62	NT	NT	103	15	6	7	03-Jun	04-Jun
164	YES	YES		YES	YES	3	25	7	62	NT	NT	103	15	6	7	04-Jun	04-Jun
181	YES	YES		YES	YES	3	25	7	62	NT	NT	103	15	6	7	04-Jun	05-Jun
187	YES	YES				NT	NT	NT	NT	NT	NT	NT	NT	NT	NT		
193	YES	YES		YES	YES	3	25	7	62	NT	NT	103	15	6	7	04-Jun	04-Jun
194	YES	YES		YES	YES	3	25	7	62	NT	NT	103	15	6	7	04-Jun	04-Jun
206	YES	YES				3	25	7	62	NT	NT	NT	NT	NT	NT	04-Jun	05-Jun
223	YES	YES				3	25 (10)	7	62 (15)	NT	NT	NT	NT	NT	NT	10-Jun	10-Jun
225	YES	YES				3	25 (10)	7	62 (15)	NT	NT	NT	NT	NT	NT	05-Jun	05-Jun
227	YES	YES				3	25 (10)	7	62 (15)	NT	NT	NT	NT	NT	NT	04-Jun	04-Jun
232	YES	YES				3	25	7	62	NT	NT	NT	NT	NT	NT	05-Jun	05-Jun
238	YES	YES		YES	YES	3	25	7	62	NT	NT	103	15	6	7	04-Jun	05-Jun
245	YES	YES				3	25	7	62	NT	NT	NT	NT	NT	NT	04-Jun	04-Jun
262	YES	YES		YES	YES	3	25	7	62	NT	NT	103	15	6	7	04-Jun	04-Jun
268	YES	YES		YES	YES	3	25	7	62	NT	NT	103	15	6	7	03-Jun	03-Jun
279	YES	YES	YES			3	25	7	62	7		NT	NT	NT	NT	05-Jun	05-Jun
284	YES	YES		YES	YES	3	25	7	62	NT	NT	103	15	6	7	04-Jun	05-Jun
285	YES	YES	YES			3	25	7	62	7		NT	NT	NT	NT	06-Jun	06-Jun
289	YES	YES		YES		3	25	7	62	7		103	15	6	7	06-Jun	06-Jun
297	YES	YES				3	25	7	62	NT	NT	NT	NT	NT	NT	04-Jun	05-Jun
299	YES	YES		YES		3	25	7	62	NT	NT	103	15	NT	NT	05-Jun	05-Jun

PLEASE NOTE: NEXT DISTRIBUTION IS ON 15TH JULY 2014

© Copyright Notice: Reports are confidential, and no data may be published without permission from UK NEQAS for H&I

UK NEQAS FOR H&I SCHEME 1A - HLA TYPING PHENOTYPING

ASSESSMENT TYPING RESULTS OF SAMPLE 1A06/2014

DESPATCHED ON 3RD JUNE

CONSENSUS HLA TYPE: HLA-

Lab No.	Supplemental DNA typing performed using:	Viability %			Testing failures	Tray Manufacturer	Comments
		U	T	B			
9	PCR-SSP		95	90		One Lambda	
11	PCR-SSP					One Lambda	
20	PCR-SSP		98	90		Own trays, One Lambda (DR)	
25	PCR-SSO		99	99		Innotrain	
28	PCR-SSO		99	99		Innotrain, One Lambda	
41	PCR-SSP	90				One Lambda	
54			95	95		One Lambda	
58	Luminex		95	90		Innotrain, One Lambda	
62			80	50			
116							NT as sample not received.
119	PCR-SSO	95		95		BioRad, One Lambda	
120	SBT	100		100		Ingen (One Lambda)	
123						One Lambda	
133	PCR-SSO	93	93			Innotrain	
139	PCR-SSO		99	99		One Lambda	
142						BioRad	
145	PCR-SSP/SSO		100	10	B-cells viability too low for typing	BioRad, Innotrain	
149						Biotest	
154			95	85			
159	PCR-SSP		90	90		One Lambda	
164			90	80		One Lambda	
181	PCR-SSP		100	100			
187							
193	PCR-SSP	98		98		One Lambda	
194	PCR-SSO		100	98		One Lambda	
206			100			One Lambda	
223	PCR-SSP	90				BioRad	
225	PCR-SSP		90			BioRad	
227	PCR-SSO	98				One Lambda	
232	PCR-SSO		90			Biotest	
238	PCR-SSO	95	95	80		One Lambda	
245	PCR-SSP		100			One Lambda	
262	PCR-SSO	95				One Lambda	
268	PCR-SSO		96	96		One Lambda	
279			100			BioRad	
284	PCR-SSO		90	90		One Lambda	
285			100				
289			98	98		One Lambda	
297	PCR-SSO	80				Ingen	
299	PCR-SSP	100		80		BioRad	