

Luminex method variability in HLA antibody specificity testing for UK NEQAS for H&I samples



Welsh Blood Service
Gwasanaeth Gwaed Cymru



for HISTOCOMPATIBILITY and IMMUNOGENETICS
UNITED KINGDOM EXTERNAL QUALITY ASSESSMENT SERVICE

Deborah Singleton, Susan Corbin & Chris Darke

WELSH TRANSPLANTATION AND IMMUNOGENETICS LABORATORY

Introduction

The UK National External Quality Assessment Service for Histocompatibility and Immunogenetics (UK NEQAS for H&I) have operated an HLA-specificity testing scheme since 1988.

In 2013 98% of participants tested samples using Luminex technology. We have noted considerable variation in reported HLA specificities despite using Luminex kits from the same manufacturer

Here we report findings from Scheme 3 – 'HLA-Antibody Specificity Analysis', for the last 5 samples distributed in 2013.

Scheme material

10 serum samples known to contain HLA antibodies are provided in 2 batches of 5 samples each year. Specificities were assessed if >74% of participants reported their presence and were considered absent if assigned by <6% of labs. Specificities reported by 6-74% participants were not assessed.

Results

All 66 participants (bar 1) tested the last 5 samples in 2013 using Luminex technology. 47 (71%) laboratories reported results based on Luminex testing only.

From the two available Luminex kit manufactures 45 participants used LABScreen (One Lambda), 13 LIFECODES (Immucor) and 7 used both.

LABScreen only users

- 6 different combinations of Mixed, PRA and Single Antigen bead (SAB) kits were used. 25 (56%) participants used SAB kits only, 13 (29%) used Mixed and SAB, 3 (7%) used Mixed, PRA and SAB, 3 (7%) used PRA and SA and 3 (7%) participants did not use SAB kits (PRA ± Mixed)
- The volume of beads used varied from 1.8-5µl (table 1)
- All participants used HLA Fusion Software but the version varied from 2.0, 3.0 and 3.2
- Of the 42 participants that completed lot information 36 used Class I SAB kit lot 008 and 6 used lot 007. All used lot 009 for Class II SAB.

Table 1: LABScreen SAB volume and MFI cut off values

Bead Volume (µl)	Number of participants	MFI Cut Off	Number of participants
1.8	1	Variable (60-2685)	6
2.0	6	500	6
2.5	11	1000	19
3.0	5	1500	2
5.0	13	2000	3

- SAB cut off values ranged from 60-2000 (table 1). 6 participants varied their cut off value for each sample.
- There was considerable variation in the SAB positive and negative control values reported for each sample (table 2).

Table 2: LABScreen SAB Control Bead Values

Sample	Positive Control Range (mean)	Negative Control Range (mean)	Pos/Neg Ratio Range (mean)
306	3118-17950 (12243)	23-617 (80)	29.1 – 695.9 (282.2)
307	3228-18695 (12283)	15-362 (42)	51.6 – 928.7 (378.8)
308	913-19200 (11418)	19-941 (195)	8.5 – 837.9 (176.8)
309	1917-17272 (11661)	20-589 (58)	20.4 – 660.9 (333.5)
310	3086-17420 (12280)	30-363 (71)	47.1 – 513.3 (222.8)

38 labs had reported full information on the bead volume, kit lot and cut-off value used for LABScreen SAB testing. 8 used the same volume of beads (2.5 µl), kit lots (008, 009) and cut-off value (1000 MFI). These were compared to the remaining 30 labs with differing test variables.

The number of non-assessed Class I specificities decreased from 55 to 38, suggesting more consistent results in the group using identical test variables (table 3).

Table 3: Non-assessed Class I Specificities

Sample	Number of non-assessed Class I Specificities (specificities reported by 6-74% of participants)		
	Differing SAB test variables (n=30)	Consistent SAB test variables (n=8)	All participants (n=65)
306	14	8	50
307	3	6	4
308	6	3	6
309	8	5	15
310	24	16	26
Total	55	38	101

Comment

These recent UK NEQAS for H&I Scheme 3's findings highlights Luminex testing variability between different laboratories using the same manufacturer's kit. These differences are probably contributing to the inter-laboratory variation seen with Luminex specificity assignments.

Further information

Full information on all UK NEQAS for H&I schemes is available at www.neqashandi.org or contact the Scheme Manager - Deborah Singleton

E-mail: ukneqashandi@wbs.wales.nhs.uk